

EU Declaration of Conformity

issued under the sole responsibility of the manufacturer

The object of this declaration: Temperature sensors type **CT...**, **CT X...**, **CT CL**Manufacturer: **APLISENS S.A.,
ul. Morelowa 7,
03-192 Warszawa**

We hereby declare under the sole responsibility, that the object of the declaration defined above comply with relevant Union harmonization legislation.

Temperature sensors of type **CT...**, **CT X...**, **CT CL** in Exi version, also with the head-mounted temperature transmitter, comply with directive:• **ATEX – 2014/34/EU** dated 26 February 2014Temperature sensors of type **CT...**, **CT X...**, **CT CL** in Exi versions, also with the head-mounted temperature transmitter, are marked with the following certificate marking:I M1 Ex ia I Ma or
II 1/2G Ex ia IIC T6...T1 Ga/Gb
II 1D Ex ia IIIC T75°C Da
KDB 06 ATEX 260X

Certificate (Module B) was issued by Notified Body no.1453, Central Mining Institute, Plac Gwarków 1, 40-166 Katowice.

The following standards were applied: EN IEC 60079-0:2018, EN 60079-11:2012, EN 50303:2000
Notification of quality assurance is issued (Module D) by Notified Body no.1453, Central Mining Institute, Plac Gwarków 1, 40-166 KatowiceTemperature sensors of type **CT...**, **CT X...**, **CT CL** with the head-mounted temperature transmitter type AT-2, ATX-2, GI-22-2, GIX-22-2 or LI-24G, LI-24G Safety comply with directives:• **EMC – 2014/30/EU** dated 26 February 2014

Conformity assessment procedure: module A. The following standards were applied EN 61326-1:2013, EN 61326-2-3:2013.

• **RoHS – 2011/65/EU** dated 08 June 2011

Conformity assessment procedure: module A, according to Decision No 768/2008/EC of the European Parliament and of the Council.

The following standard was applied: EN IEC 63000:2018.

Temperature sensors of type: **CT...**, **CT X...**, **CT CL** in all versions comply with the requirements of standards: EN 60584-1:2013, EN 60751:2008

Warsaw, 01.07.2021


Adam Żurawski
General Manager